

**AGROLAB LUFA** Dr.-Hell-Str. 6, 24107 Kiel

Date 08.12.2023  
Customer no. 10083246

**REPORT**

Order **3350297** Order no: 2373  
Sample no. **198487**  
Sample acceptance **04.12.2023**  
Date of sampling **no information**  
Sample taker **Client**  
Customer sample description **sample 54:  
Hericium Glucan extract  
Lotnumber: HEE-G-23101601  
Ident.-Nr.: 100024**

Packaging **1x plastic bag, à 50 g**

Unit	Result	Limit value	Substance	Method
------	--------	-------------	-----------	--------

**Trace elements / Heavy metals / Halogenides**

Inorganic arsenic	mg/kg	<b>0,095</b>	OM	DIN EN 16802 : 2016-07
-------------------	-------	--------------	----	------------------------

*Parameter-specific analytical measurement uncertainties and information regarding the method of calculation will be provided upon request if the reported results are above the parameter-specific limit of quantification.*

*Explanation: OM = on original matter; DM = on dry matter base*

*Start of testing: 05.12.2023  
End of testing: 08.12.2023*

*The results are related only to the samples tested. In cases where the laboratory has not been responsible for sampling, the reported results apply to the samples as received. The laboratory is not responsible for the information provided by the customer. The customer information, if any, presented in this test report is not subject to the accreditation of the laboratory and may affect the validity of the test results. Duplication of this document or of parts of it requires the authorization from laboratory. In accordance our agreement in writing in the order confirmation, the results in this test report are in a simplified form in the context of DIN EN ISO/IEC 17025:2018, paragraph 7.8.1.3. In conformity assessment, the economic approach is used as the decision rule (a non-conformity exists if the measurement result is included measurement uncertainty above the specification or standard), as long as nothing else has been determined by corresponding legal or normative bases.*



**AGROLAB LUFA Service-Team L1, Tel. 0431/1228-337**  
**Group leader: Wiebke Stelzer**  
**Food chemist/counter-sampling expert**

The activities reported in this document are accredited according to DIN EN ISO/IEC 17025:2018. Only not accredited activities are identified by the symbol "(\*)".